

PVC COLOURING COMPOUNDING & PROCESSING

64, Phase-I, Opp.Citizen Industries, G.I.D.C Naroda, Ahmedabad-382330

Phone : +91-79-2281 2004, Fax : +91-79-2282 2006, E-Mail : pvcl@pvclplastics.com, Web : <http://www.pvclplastics.com>

Product Specification for Urine Drainage Tubing

Type	Grade	Physical Properties		Application
		Specific Gravity	Hardness Shore 'A' 15 Sec @ 23 ⁰ C	
Test Method →		ASTM D 792	ASTM D 2240	
UT-01	ENH-UT-65	1.19	65	Urine Drainage Tubing
	ENH-UT-68	1.20	68	
	ENH-UT-70	1.21	70	
	ENH-UT-72	1.21	72	
	ENH-UT-74	1.22	74	
	ENH-UT-80	1.22	80	
	ENH-UT-84	1.23	84	
	ENH-UT-86	1.24	86	
	ENH-UT-88	1.25	88	

Note: The information given here in is based on the present state of our knowledge but any conclusions and recommendations are made without liability on our part. Buyers and users should make their own assessment of our products under their own conditions and for their own requirements.